

Doc Title:	MP.MP+_Declaration Of Conformity MDD						
Doc #:	ML-PR-34026	Rev.:	4	Page:	1 of 1		Status:
Type:							Unclassified Uncontrolled when printed or copied!
Created by:	Sasa Karpeti		Ownership:	Regulatory Affairs			
Eff. Date:	Jul 6, 2020	Origin:	Internal	Owner:	Sasa Karpeti		
Item Name	MP.MP+ Model						
Item Code:	34000	Item Version:	6	Item ERP No.:	N/A	Item Level:	Model
Governance	Doc #:	Title				Revision	Released
Used Form	ML-TMP-00001	NA				NA	Jul 5, 2018
Governed WIN	ML-TMP-00001	NA				NA	Jul 5, 2018
Governed SOP	ML-TMP-00001	NA				NA	Jul 5, 2018
Filename:	ML-PR-34026_MP.MP+_Declaration Of Conformity MDD_Rev.4.docx					Copy created:	2020-07-01

1 Document Approval - Electronic Signatures

Christos Freris;Document Control Review;Friday, June 5, 2020 11:41:40 AM EEST
 Achilles Tsoukalis;Review;Wednesday, June 17, 2020 6:50:15 PM EEST
 Sasa Karpeti;Final Approval;Thursday, June 18, 2020 9:03:33 AM EEST
 Christos Freris;Final Approval;Monday, June 22, 2020 12:32:13 PM EEST

2 Declaration of Conformity

We,

the medical device manufacturer,

Micrel Medical Devices S.A.
42 Konstantinoupoleos Str.
Koropi/Athens GR-19441
GREECE

declare under our sole responsibility

that the products listed in the document

"ML-PR-34014_MP.MP+_Registration_Item List"

- are in conformity with the essential requirements and principles for safety and performance of the
Medical Devices Directive 93/42/EEC ("MDD") amended by Council Directive 2007/47/EC
- are classified Class IIb according to the Annex IX, rule 11, of the above directive
- are CE certified according to the conformity assessment and supervision of the

Notified Body SGS Belgium NV (CE1639),
SGS House, Noorderlaan 87, 2030 Antwerp, Belgium

July 5 2020
Kopium
Sasa Karpeti Q.A.M.

This declaration is signed electronically and valid until 2024-05-24,
 the expiry date of the EC Full Quality Assurance System Certificate BG19/871877.

Doc Title:	MP.MP+_Registration_Item List					
Doc #:	ML-PR-34014	Rev.:	1	Page:	1 of 1	Status: Final Approval
Type:	Item List					Unclassified Uncontrolled when printed or copied!
Created by:	Sasa Karpeti		Ownership:	Regulatory Affairs		
Eff. Date:		Origin:	Internal	Owner:	Sasa Karpeti	
Item Name	MP.MP+ Model					
Item Code:	34000	Item Version:	6	Item ERP No.:	N/A	Item Level: Model
Governance	Doc #:	Title		Revision	Released	
Used Form	ML-TMP-00001	NA		NA	Jul 5, 2018	
Governed WIN	ML-TMP-00001	NA		NA	Jul 5, 2018	
Governed SOP	ML-TMP-00001	NA		NA	Jul 5, 2018	
Filename:	ML-PR-34014_MP.MP+_Registration_Item List_Rev.1.docx				Copy created:	2019-09-20

1 Document Approval - Electronic Signatures

Christos Freris;Document Control Review;Friday, August 9, 2019 4:06:38 PM EEST

Achilleas Tsoukalis;Review;Tuesday, August 13, 2019 4:35:34 PM EEST

2 Sales Packs

Pos.	Item Ref. Nr.	Item Name
1	KP1.12.201.4	MP101+
2	KP1.12.202.4	MPmlh+ BD-20
3	KP1.12.203.4	MPDaily+
4	KP1.12.206.4	MPmlh+ TERRUMO-60cc
5	KP1.12.207.4	MPmlh+ LMT-20
6	KP1.12.208.4	MPmlh+ CODAN-20
7	KP1.12.209.4	MPmlh+ BBRAUN OMNIFIX 20cc
8	KP1.12.210.1	MPmlh+ TERRUMO/TYCO MONOJECT 20cc
9	KP1.12.220.6	MPmlh+ (Multi Syringe)
10	KP1.12.213.4	MPmlh+ LMT-20 EVER
11	KP1.12.215.3	MPmlh+ BD-20 EVER
12	KP1.12.217.3	MPmlh+ CODAN-20 EVER
13	KP1.12.222.1	MPmlh+ (Multi Syringe) EVER
14	KP1.12.212.4	MP THALAPUMP-20

3 Accessories

Pos.	Item Ref. Nr.	Item Name
1	KM1.YY.604.1	MP+ Carrying bag with transparent window 20-50ml
2	KS1.12.164.1	MP+ Soft Locker box 20 ml
3	KS1.12.165.1	MP+ Soft Locker box 50 ml
4	KM1.EE.121.1	MP+ Pole Clamp
5	KM1.YY.561.1	Disposable bag for MP+ pumps