




NOVAERUS

MDD 93/42/EEC Declaration of Conformity

MANUFACTURER:	Novaerus Ireland, Ltd. DCU Innovation Campus, Old Finglas Road Glasnevin, Dublin 11, Ireland Phone: +35319072750
PRODUCTS:	Novaerus Infection Control Model NV800 Air Disinfection Unit
CLASSIFICATION & ANNEX:	Class I under Medical Device Directive (MDD) 93/42/EEC Rule 1 and Rule 12. The conformity assessment procedure per Article 11 for a Class I device is Annex II of the MDD 93/42/EEC.
DECLARATION:	We herewith declare that the above-mentioned products meet the provisions of the Medical Device Directive (MDD) 93/42/EEC for medical devices. All supporting documentation is retained under the premises of the manufacturer.
STANDARDS APPLIED:	<ul style="list-style-type: none">• ISO 14971: Medical Devices: Application of Risk Management to Medical Devices• IEC 60601: Medical electrical equipment<ul style="list-style-type: none">• Part 1: General requirements for basic safety and essential performance;• Part 1-2: Collateral standard: Electromagnetic compatibility – requirements and tests• ISTA Procedure 2A Partial-Simulation Performance Test Procedure: Packaged Products 150lb (68 kg) or Less• UL 867: UL Standard for Safety for Electrostatic Air Cleaners, Section 40, Ozone Test, Fifth Edition• ISO 15223-1: Medical devices: Symbols to be used with medical device labels, labelling and information supplied – Part 1: General requirements.
DATE OF ISSUE:	23 April 2020
SIGNATURE:	 _____ Declan Kiely, Quality Director, Novaerus Ireland Ltd.